INTOX EC/IR II

Quality Assurance Worksheet



Instrument Serial Number

010528

Worksheet Start Date

12/29/2014

Location

Fairfax County Adult Detention Center

Address

10520 Judicial Dr., Fairfax, VA 22030

DFS Technician

Heather Stanton

License No.

18910

Laboratory

V On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG)

Reference Barometer (mm HG)

755

Reference Barometer(RB)Serial # 009113

RB Calibration Due

V

4/28/2015

Measurement Assurance Check

Standard (sea level)

PA Target

PA Target

minimum

maximum

Sample 1 Sample 2

Precision

-0.003sample min sample max

0.003 0.000

Sample 3

0

0.000

maximum

Sample 1

Precision

Standard (sea level)

minimum -0.003 sample min

0.003 sample max

Sample 2

0

0.000

0.000

Sample 3

Dry gas standard Lot No. (with tank no.)

0.300

0.100

AG419702-50

Replaced dry gas standard (+O-ring)

Supplies Mouthpieces

Installed at Location

Removed to DFS-Central


Certificates of Analysis

Operator Worksheet

Other:



Notes:



Instrument	Serial	Number
------------	--------	--------



Certification Date



Calibrated

Certified

✓ Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level) 0.250 Precision 0	PA Target	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.150 Precision 0	PA Target	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.080 Precision 0	PA Target	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.020 Precision 0	PA Target	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Date

12/30/14

Issuing Analyst

Date

12/30/14

1910





COMMONWEALTH OF VIRGINIA DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

1		NAME OF COURT	
INSTRUMENT, TEST,		DFS	
BREATH ANALYSIS			
SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY	
STANTON, HEATHER, M		DFS Central Lab	
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED	
18910	10/01/2016	12/29/2014	
TEST EQUIPMENT NUMBER		於[1] [2] [2] [2] [2] [2] [2] [2] [2] [2] [2	
010528		ない。これは、これは、これは、これは、これが、これは、これがない。	
RESULIS: TIME SAMPLE TAKEN	11:43 EST	GRAMS PER 210 LITERS OF BR	EATH
ATTEST:			
I CERTIFY THAT THE ABOVE IS AN ACCURATE EQUIPMENT AND IN ACCORDANCE WITH THE CONDUCTED IN ACCORDANCE WITH THE TEST THE ACCUSED WAS ADVISED OF HIS MENT USED TO PERFORM THE BREATH TEST, A	E METHODS APPROVED THE DEPARTMENT'S SRIGHT TO OBSERVE TH	CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TO BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY	ST WAS ON OF EQUIP-
BREATH TEST OPERATOR			
☐ I HAVE RECEIVED A COPY OF THIS		ANALYSISSUBJECT'S SIGNATURE TE OF ANALYSIS	



IntoxNet MIS Report Report Generated 30 Dec 2014 at 08:21



Test Results

Test Status Success

	Instrument Serial N	umber 010528
Test # 001461 Subject Test		
Test Location 1 Department of	Test Location 2 Forensic Sci-	ience Test Location 3
Test Date 29 Dec 2014	Test Time 11:	37 System Check Passed
	Remote/Local Loc	cal
Operator's Last Name STANTON	Operator's First Name HEA	THER Operator's Middle Initial M
Agency DFS Central Lab		License Number 18910
Card Serial Number 118910	Effective Date 10/01	/2014 Expiration Date 10/01/2016
Subject's Last Name INSTRUMENT		Subject's First Name TEST
. Subject's Middle Initial	Subject's Date of Birth 0	00/00/0000 Subject's Sex Male
Driver's License Number		Driver's License Expiration 00/00/0000
Driver's License State		Court Name DFS
End Date 29 Dec 2014 End Time 1	1:44 Result Time 11	1:43 Result Date 29 Dec 2014 Result 0.00
Data Type DIAG	Sample Value Pass	Sample Time 11:37
Data Type BLK	Sample Value 0.00	OO Sample Time 11:38
Data Type CHK	Sample Value 0.10	OO Sample Time 11:39
Data Type BLK	Sample Value 0.00	OO Sample Time 11:40
Data Type SUBJ	Sample Value 0.00	Sample Time 11:41
Data Type BLK	Sample Value 0.00	OO Sample Time 11:42
Data Type SUBJ	Sample Value 0.00	•
Data Type BLK	Sample Value 0.00	•
Standard Type Dry Gas Std		Standard Value 0.099
Standard Lot Number AG419702-50		Standard Expiration Date 07/16/2016
Tank Pressure 1022		Barometric Pressure 753 mmHg
• • • • •		Blow Volume 1945 cc End-of-Blow Time 11:41
Blow Sample Number 2 Blow D		Blow Volume 1768 cc End-of-Blow Time 11:43
Tamper Evident Stamp 6240	5a85	Test Status Code 0